






# DENTAL CITY EARLOOP MASKS ASTM LEVEL 2



**TO WHOME IT MAY CONCERN: AS OF DECEMBER 15, 2017 DENTAL CITY EARLOOP MASKS ARE CLASSIFIED AS MEETING ASTM LEVEL 2 REQUIREMENTS. *PLEASE NOTE: DENTAL CITY MASK PACKAGING STILL REFLECTS AN ASTM LEVEL 1 RATING. THIS WILL BE UPDATED AS SOON AS POSSIBLE.***

## **ASTM F2100 Mask Performance Rating:**

The ASTM F2100-11 Standard Specification for Performance of Materials used in Medical and Dental Face Masks assists in selecting the appropriate mask design, fit and filtration that best matches the protection needs for individual procedures or risk levels.

## **ASTM Level 2 Dental Earloop Masks Requirements**

|  |         |
|--|---------|
| Fluid Resistance, mm Hg                      | 120     |
| BFE  | ≥98%    |
| PFE, @ 0.1 micron                            | ≥98%    |
| Delta P, mm H <sub>2</sub> O/cm <sup>3</sup> | <5.0    |
| Flame Spread                                 | Class 1 |

## **The following masks are rated as ASTM Level 2:**

| <u>SKU</u> | <u>SKU Name</u>                           |
|------------|---|
| 25-205BL   | Dental City Earloop Masks Blue 50/box     |
| 25-205GR   | Dental City Earloop Masks Green 50/box    |
| 25-205LV   | Dental City Earloop Masks Lavender 50/box |
| 25-205PK   | Dental City Earloop Masks Pink 50/box     |
| 25-205WH   | Dental City Earloop Masks White 50/box    |