


## Safety Data Sheet

Trade Name: **Genesis Temp Crown and Bridge**

1.0 Group																
1.1	Commercial product name / designation <b>Genesis Temp Crown and Bridge</b>															
1.2	Application / Use Dental material for use by dental professional only.															
1.2.2	SIC 851 Human health activity															
1.2.3	Use Category 55															
1.3	Manufactured for <b>The Pelican Group</b> 1195 Airport Road Lakewood, NJ 08701 USA Telephone: 1 732 666-9090 / Fax: 1 732 666-9091 Email : sales@thepelicanusa.com															
1.4	Emergency Telephone Number 1-732-666-9090															
2.0 Hazards Identification																
2.1	Classification															
2.1.1	Classification according to Regulation (EC) No 1272/2008 [CLP]															
	<table border="1"> <thead> <tr> <th>Hazard Class</th> <th>Hazard Category</th> <th>Hazard Statement</th> </tr> </thead> <tbody> <tr> <td>Eye irritation</td> <td>2</td> <td>H319</td> </tr> <tr> <td>STOT SE</td> <td>3</td> <td>H335</td> </tr> <tr> <td>Skin irritation</td> <td>2</td> <td>H315</td> </tr> <tr> <td>Skin sensitization</td> <td>1</td> <td>H317</td> </tr> </tbody> </table>	Hazard Class	Hazard Category	Hazard Statement	Eye irritation	2	H319	STOT SE	3	H335	Skin irritation	2	H315	Skin sensitization	1	H317
Hazard Class	Hazard Category	Hazard Statement														
Eye irritation	2	H319														
STOT SE	3	H335														
Skin irritation	2	H315														
Skin sensitization	1	H317														
2.1.2	Classification according to Directive 67/548/EEC Irritant; Xi; R 36/37/38 – 43 (See SECTION 16 for full text of risk phrases)															
2.2	<b>GHS Label Elements</b>															
	Hazard Pictograms															
																
	Signal Word: <b>WARNING</b>															
	<b>Restricted to use by dental professional only.</b>															
	<b>Hazard Statements</b>															
	H319: Eye irritation. 2. May cause eye irritation.															
	H335: STOT SE. 3. May cause respiratory irritation.															
	H315: Skin irritation. 2. May cause skin irritation.															
	H317: Sensitization. 1. May cause an allergic skin reaction.															
	<b>Precautionary Statements</b>															
	P261: Avoid breathing vapor.															
	P280: Wear protective gloves and eye protection															
	P305+P351: If in eyes, rinse cautiously with water for several minutes.															
	P337+P313: If eye irritation persists, get medical advice/attention.															
	P302+P352: If on skin, wash with plenty of soap and water.															
	P333+P313: If irritation or rash occurs, get medical advice/attention.															
	P410+P411: Protect from sunlight. Store at temperature not exceeding 27°C / 80°F.															
3.0 Composition																
3.1	Chemical characterization of the preparation: Glass-filled methacrylate resins in a two-part, dual-cure paste.															
3.2	Hazardous ingredients															
	<table border="1"> <thead> <tr> <th>CAS Number</th> <th>Name of the Ingredient</th> <th>Concentration</th> <th>Classification per 67/548/EEC</th> <th>Classification per Regulation (EC) No.1272/2008 (CLP).</th> </tr> </thead> <tbody> </tbody> </table>	CAS Number	Name of the Ingredient	Concentration	Classification per 67/548/EEC	Classification per Regulation (EC) No.1272/2008 (CLP).										
CAS Number	Name of the Ingredient	Concentration	Classification per 67/548/EEC	Classification per Regulation (EC) No.1272/2008 (CLP).												

**Safety Data Sheet**Trade Name: **Genesis Temp Crown and Bridge**

Proprietary	Uncured methacrylate resins	55-60%	Xi (Irritant), R 36/37/38 - 43	Eye irritation, 2, H319 STOT SE 3, H335 Skin irritation, 2, H315 Skin sensitization, 1, H317
112945-52-5	Silica, amorphous	5 %	Xi (Irritant), R 36/37/38	Eye irritation, 2, H319 STOT SE 3, H335 Skin irritation, 2, H315

**4.0 First Aid Measures**

4.1	General Information	May be irritating to eyes, mucous membranes or skin on contact. Avoid contact with eyes and skin. May cause sensitization by skin contact. Show this safety data sheet to medical personnel. Get medical attention in case of uncertainty.
4.2	Inhalation	Move to fresh air. If necessary, administer oxygen and/or artificial respiration. Seek medical attention.
4.3	Skin Contact	Take off contaminated clothing. Wash skin thoroughly with soap and water.
4.4	Eye Contact	Keep eyelids apart and flush with running water for 15+ minutes. Get medical attention.
4.5	Ingestion	Rinse mouth with water. Do not induce vomiting. Get immediate medical attention. May be irritating to mucous membranes. Never give anything by mouth to an unconscious person.
4.6	Precautions for first responders	Ventilate the area. Wear eye and skin protection.
4.7	Information for physicians	
	Symptoms	Irritation or redness in eyes, throat or on skin.
	Hazards	May be irritating to eyes, mucous membranes and skin. May cause sensitization by skin contact.
	Treatment	As above under First Aid.

**5.0 Fire Fighting Measures**

5.1	Suitable extinguishing media	Carbon dioxide, dry chemical, alcohol foam, or water fog. Water spray may be used to keep fire exposed containers cool.
5.2	Extinguishing media to avoid	Do not use direct water stream
5.3	Special exposure hazards in a fire	Heat may cause polymerization with rapid release of energy.
5.4	Special protective equipment for fire-fighters	Self-contained breathing apparatus

**6.0 Accidental Release Measures**

6.1	Personal precautions.	Wear safety glasses, gloves and lab coat.
6.2	Environmental precautions	Avoid releasing large amounts of uncured material into the environment Cured / set-up material is, to our knowledge, inert.
6.3	Method for clean up	Contain spilled material. Absorb or wipe up spill paper towels or cloths. Collect for disposal in a covered container. Wash area of spill with alcohol or soap and water.

## Safety Data Sheet

Trade Name: **Genesis Temp Crown and Bridge**

<b>7.0</b>	<b>Handling and Storage</b>	
7.1	Handling	For use only by dental professionals. Follow good hygiene practices. Avoid direct, strong light sources, temperature extremes (>27°C/80°F, <5°C/40°F), contamination, cross contamination. Recap immediately after use.
7.2	Storage	Store tightly capped in original container at cool room temperature (<25°C). Avoid direct, strong light, sources of ignition, extremes of temperature. Shelf life for unopened product is two years from date of manufacture, provided that the material has been stored properly.
7.3	Specific uses	Provisional dental material
<b>8.0</b>	<b>Exposure Controls / Personal Protection</b>	
8.1	Exposure limit values	PEL: Not established. TLV: Not established.
8.2	Exposure controls	
8.2.1	Occupational exposure controls	No special equipment required under normal conditions of use of this product in the quantity provided.
8.2.1.1	Respiratory protection	Good general ventilation is sufficient to control any airborne.
8.2.1.2	Hand protection	Surgical gloves
8.2.1.3	Eye protection	Safety glasses
8.2.1.4	Skin protection	Good personal hygiene and safety practices; wear a lab coat.
8.2.1.5	Other controls	Emergency eye wash fountain. Wash hands after use.
8.2.2	Environmental exposure controls	Cure material before disposing.
<b>9.0</b>	<b>Physical and Chemical Properties</b>	
9.1	Characteristics	
9.1.1	Appearance / Color / Physical state	Two pastes: tooth-colored base; off-white catalyst
9.1.2	Odor	Faint, characteristic
9.2	Important health, safety and environmental information	
9.2.1	pH	Not determined
9.2.2	Boiling point	Not determined
9.2.3	Flash point	Not determined
9.2.4	Flammability (solid, gas)	Not applicable
9.2.5	Explosive properties	Not applicable
9.2.6	Oxidizing properties	Not determined
9.2.7	Vapor pressure	< 1 mm Hg
9.2.8	Specific gravity	1.520
9.2.9	Solubility in water	Nil
9.2.10	Partition coefficient	Not determined
9.2.11	Viscosity	Not determined

## Safety Data Sheet

Trade Name: **Genesis Temp Crown and Bridge**

9.2.12	Vapor density	Not determined
9.2.13	Evaporation rate	Not determined
<b>10.0</b>	<b>Stability and reactivity</b>	
10.1	Conditions to avoid	Temperature > 27°C / 80°F, light, cross-contamination of parts.
10.2	Materials to avoid	Reducing and oxidizing agents, peroxides, amines.
10.3	Hazardous decomposition products	Under fire conditions and with amounts greater than that supplied in this product, hazardous polymerization may occur with heat build-up and release of carbon monoxide, carbon dioxide, and oxides of nitrogen.
10.4	Further information	Stable material if stored and used as directed. Polymerization will occur when light-sensitive material is exposed to direct light.
<b>11.0</b>	<b>Toxicological information</b>	
11.1	Acute toxicity	Not toxic. Minimal health hazard in the quantities present in this product and under normal conditions of use.
11.2	Irritation and corrosiveness	May be irritating to eyes, mucous membranes or skin on contact or with prolonged exposure.
11.3	Sensitization	May be sensitizing. Prolonged or frequent skin contact may cause allergic skin reactions in some susceptible individuals.
11.4	Sub-acute, sub-chronic and prolonged toxicity	Prolonged and/or frequent skin contact may cause allergic skin reactions in susceptible individuals. Prolonged exposure to large amounts (more than in this product) may cause eye and respiratory irritation.
11.5	Carcinogenicity, Mutagenicity, Reproductive Toxicity	None known
11.6	Empirical data	Testing confirmed biocompatibility of Tuff-Temp Plus.
11.7	Clinical experience	Tuff-Temp Plus has been on the market for more than four years with no reports of adverse events.
<b>12.0</b>	<b>Ecological Information</b>	
12.1	Ecotoxicity	Avoid release of uncured material into the environment. To the best of our knowledge, polymerized material is inert. No other information is available. Follow all government regulations.
<b>13.0</b>	<b>Disposal Considerations</b>	
13.1	Regulations	Follow all local and national government regulations in disposing material or contaminated packaging.
<b>14.0</b>	<b>Transport Information</b>	
14.1	Restrictions	None. Not regulated.
<b>15.0</b>	<b>Regulatory Information</b>	
15.1	EU Regulations	Class IIa medical device under MDD 93/42/EEC
15.2	US FDA	Class II medical device.

**Safety Data Sheet**Trade Name: **Genesis Temp Crown and Bridge**

15.3	Health Canada	Class III medical device
<b>16.0</b>	<b>Other information</b>	
16.1	List of relevant R phrases	R36/37/38: Irritating to eyes, respiratory system and skin. R43: Sensitizing by skin contact.
16.2	Hazard Statements	H319: Eye irritation. Hazard category 2. H335: Specific Target Organ Toxicity - Single exposure; hazard category. 3. Respiratory tract irritation. H315: Skin irritation. Hazard category 2. H317: Skin Sensitization. Hazard category 1.
16.3	Precautionary Statements	P261: Avoid breathing vapor. P280: Wear protective gloves and eye protection P305 + P351: If in eyes, rinse cautiously with water for several minutes. P337 + P313: If eye irritation persists, get medical advice/attention. P302 + P352: If on skin, wash with plenty of soap and water. P333 + P313: If irritation or rash occurs, get medical advice / attention. P410 + P411: Protect from sunlight. Store at temperature not exceeding 27°C / 80°F.
16.4	Restrictions on use	To be sold to and used by dental professionals only.
16.5	Further information	The information presented herein is believed to be factual as it has been derived from the works of persons believed to be qualified experts. However, nothing contained in this information is to be taken as a warranty or representation for which The Pelican Group bears legal responsibility. The user should review any recommendations in the specific context of the intended use to determine whether they are appropriate.
16.6	Sources of key data	National Institute for Occupational Safety (NIOSH) US Occupational Safety and Health Administration (OSHA) Eur-Lex European Union Law: Regulation (EC) No. 1272/2008 (CLP) and Regulation (EC) No. 1907/2006 (REACH). Guidance on the compilation of safety data sheets. Version 1.1; December 2011. European Chemicals Agency
16.7	Information which has been added, deleted or revised.	This Safety Data Sheet has been revised to meet the requirements of the GHS SDS format, Regulations (EC) No. 1272/2008 (CLP) and (EC) No. 1907/2006 (REACH). Specifically, Sections 2.1, 2.2, 3.2, 16.2, 16.3 have been modified.